

# **User Manual**

Please read this manual thoroughly before using and following the instructions.


#### **Part Instruction**

#### ① Suction Connection

This is the part that connects Dr. Vortex and Suction.

#### 2 Protection Side

It is a protection side that prevents the spread of oral secretions. Material: PET



#### How to Use

#### Preparation before use

- 1) The product should be used according to intended use.
- 2) Always check for wear, cracks, deformation, corrosion, damaged parts before use.
- 3) Do not use the product if it is damaged.
- 4) Replace the damaged parts with spare parts immediately before use.

- 1) The product should be used after reading this user's manual.
- 2) Use a ballpoint pen to tighten the Suction Connection so that it is not damaged.

#### Storage & Maintenance after use

- 1) Do not store with heavy objects on the product.
- 2) The product is recommended as disposable use.

# How to Use: **Manual Movement**



1) Put the ballpoint pen inside Dr. Vortex to fix it. 2) To avoid breaking, push the connection part of Dr. Vortex to tighten up



# **Precautions**

1) When fastening to the suction entrance directly, the durability decreases due to the change in appearance.



- 2) Do not use the product if it is not fully punctual.
- 3) Replace the damaged part immediately if you spot it.



# Protect your health by Dr. Vortex

During surgery and scaling, **foreign substances and blood flow into the body** without knowing. **Dr. Vortex effectively protects the doctors, hygienist and patients** through the immediate blocking generated during dental procedure such as escavation in tooth.



# **During dental procedures**



Foreign substances and blood generated during treatment

### **Clinical Case**



Prevents spreading of the foreign substances and blood by using Dr. Vortex

